


RehaCom

computer-assisted cognitive rehabilitation - brain performance training



Visuo-motor Coordination

RehaCom[®]

computer-assisted cognitive rehabilitation

by HASOMED GmbH

This manual contains information about using the RehaCom therapy system.

Our therapy system RehaCom delivers tested methodologies and procedures to train brain performance .
RehaCom helps patients after stroke or brain trauma with the improvement on such important abilities like memory, attention, concentration, planning, etc.

Since 1986 we develop the therapy system progressive.
It is our aim to give you a tool which supports your work by technical competence and simple handling, to support you at clinic and practice.

HASOMED GmbH
Paul-Ecke-Str. 1
D-39114 Magdeburg
Germany

Tel. +49-391-6230112

Table of contents

Part I Description of the training	1
1 Training tasks	1
2 Performance feedback	2
3 Structure of the level of difficulty	3
4 Training parameters	5
5 Evaluation	7
Part II Theoretical Concept	7
1 Basic foundations	7
2 Aim of the training	10
3 Target groups	11
4 Bibliography	12
Index	16


1 Description of the training

1.1 Training tasks


Visuo motor co-ordination is trained by using a coloured (Rotor), which glides across the screen in various complicated [patterns of movement](#) (Picture 1). On the screen there is also a smaller (Cursor), which the patient controls by using the Reha-Com panel. The patient has to try to maintain as much contact as possible with the rotor as it glides across the screen.

The 'time outside' is measured, i.e. the total time in which the cursor is not in contact with the rotor. The measurement is calculated as a percentage, in relation to the **length of the task**. A task is considered as solved when the 'time outside' is less than or equal to the established minimum 'time outside'. A task is considered not to be solved when the 'time outside' is larger than the established maximum 'time outside'. If the time measured lies between these two previous values then the same pattern of movement is repeated. In this way the training process remains adaptive. In addition, the parameter **repetition** can also be set up to establish if a patient should advance to the next level. The error count is also registered as the number of the times the "Cursor was outside of the rotor" (cf. also [Evaluation](#)), however this is not used to evaluate a task.

as an alternative to the **abstract** rotor and cursor forms, a **concrete** form can also be used. In this situation, a beetle(cursor) must be placed on a flower(rotor). Picture 2 shows this very situation. In this way, with various types of cursors and rotors a high level of motivation is achieved. Further examples are frog on a leaf, egg in a basket and mouse on the cheese.



Picture 1. Training of rotor type *abstract*.



Picture 2. Example of Rotor type *concrete*.

1.2 Performance feedback

When the acoustic feedback is activated, then a continuous error tone is generated every time a customer makes a mistake (Cursor outside of the rotor). When the [Cursor](#) is moved back within the rotor zone then the signal is deactivated. It is advisable to use this form of feedback at all times.

If there is more than one patient working in the room then this type of acoustic feedback could cause interference. In this case the visual feedback should be used. The Rotor changes colour to red if the cursor is outside the Rotor zone.

After completing a task the patient is informed as to how they performed, the number of errors are then shown and the percentage of 'outside-time'. During the training the level can be seen on the bottom right hand side of the screen.

1.3 Structure of the level of difficulty


The nature of the [Rotors](#) is the key to the levels of difficulty:

- . The pattern of movement (32 different patterns of movement - see Picture 3),
- . The size (3 rotor sizes - large, medium and small) and
- . The speed (3 rotor speeds - slow, medium and fast)

The patterns of movement are divided into 5 sub groups:

- . simple linear movement (patterns 1-2),
- . combined linear movement (patterns 3-14),
- . curved movement (patterns 15-24),
- . combined curved and straight movements (patterns 25-31) and
- . a pseudo stochastic movement (Muster 32).

In the last case - pattern 32 it is impossible to predict the nature of the movement.



Picture 3. Patterns of movement.

The patterns of movement are formulated in such a way as to facilitate an increasing complexity of movement with individual pattern number. The pattern numbers which have been marked are the ones which will be used in the procedure. If they are not marked then the patient will skip this pattern during the training. In this way, the patient can concentrate on certain movements while excluding others.

During the training the patient works through the patterns in sequence. When working with a particular pattern of movement, either the size or the rate of the rotor (speed) can be varied - in each case by 3 stages. One chooses the size or speed by using the control parameter. If the control parameter is set at speed, one can then choose the size of the rotor by selecting the parameter size - large, medium or small (stages 1-3). It will then remain the same at each individual task. If the control parameter is set to size, the speed can then be modified (stages 1-3 slow, medium and fast). In this way, target-orientated training is possible. The tasks are more difficultly when the speed is set at fast or the size set to small.


In order to alternate between the patterns of movement, the control modes constant and sequential are available. In the mode constant - each pattern is trained in 3 stages (slow, medium, fast or large, medium, small). After the patient has successfully completed pattern number 1, at all 3 stages, he then moves on to pattern number 2. In the mode sequential - at first all the patterns in stage 1

should be trained. Only when the most difficult pattern is mastered in stage 1, can the patient start work on the patterns at stage 2.

The performance graphic at the end of training displays the level of difficulty in relation to the solved patterns. The maximum level of difficulty is 96 (32 patterns of movement * 3 possible variations (speed or size)).

1.4 Training parameters

In the **RehaCom basic foundations**, some general information (references) is given on the Training Parameters and their properties. This information (references) should be taken into further consideration. Picture 4 shows the parameter-menu.



Picture 4. Parameter-Menu.

Duration of training/Cons. in mins:

A training period of 20-30 minutes is recommended.

Duration of task in seconds:

The duration of the task should be clearly established (recommended 60-180 s).

After this the task is ended and evaluated.

Minimum and maximum time outside:

See [Training tasks](#).

Repetition:

The patient starts new task only if the current task was solved in accordance with established repetitions. The level of difficulty is reduced only when the can't solve a task in accordance with the number of repetitions. In this way the changing of levels is directly linked to the patient abilities.

Control parameter, Size and Tempo:

See [Structure of the level of difficulty](#).

Control mode:

See [Structure of the level of difficulty](#).

Rotor type:

See [Training tasks](#).

Orientation:

The training field can be set up to the left or the right of the screen. this option is of interest to patients who suffer from [Neglect](#) or some other type of disturbance to the field of vision.

Feedback:

See [performance feedback](#).

With each new set up of the training the following defaults are automatically installed (standard level):

Duration of training	20 Minutes
Duration of task	180 Seconds
Minimum time outside	10 %
Maximum time outside	35 %
Repetition	1
Rotor type	concrete
Control parameter	large
Tempo	slow
Control mode	constant
Orientation	left
Feedback acoustic	on
Feedback visual	off

1.5 Evaluation

The diverse possibilities of data analysis for the determination of the further training strategy are described in the **RehaCom basic foundations**.

In the pictures as well as the tables, alongside the setting for the [trainings parameter](#), the following information is available:

Level	current level of difficulty
Training time (effective)	effective Training time
Pauses	Number of pause by the patient
Figure	Number of patterns of movement
Size	large / medium / small
Tempo	slow / medium / fast
Number (Out)	Number of times the Cursor was outside the Rotor
Time-outside	Time-outside in %

In this way it is possible to give the patient advise on their short-comings.

2 Theoretical Concept

2.1 Basic foundations

Most human movement functions on the basis of **co-ordination** of several central **motor**, **visual** and **proprioceptive** systems. In the case of numerous activities, like the grabbing of objects, writing or playing a music instrument, an exact co-ordination of eyes, head and hands is required. During the act of moving (motor action), visual control has an essential role to play - in particular, when using tools and when learning new sequences of movement. (*performance of visuo-motor transformation* cp. [Guedon et al.](#), 1998).

Learning to produce motion

During voluntary movements, in particular with tool use, alongside fully functional motor skills, an intact visual system as well as an intact proprioception is necessary (cp. [Guedon et al.](#), 1998). The position of the hand and the target are probably represented as visual "cards", proprioceptive and motor "rooms", where visual and motor representations correspond to each other. (Bedford, 1994; cp. [Guedon et al.](#), 1998). If one of these systems experiences problems, motoric control can be compensated, up to a certain degree, by the other systems. (cp. [Guedon et al.](#), 1998).

In the case of the adaptive control of movements, the proprioceptive part is more important to perception, if one learns under disproportional feedback conditions, as for example, in the case of the distorted visual response through mirrors or prisms. The inconsistency between proprioceptive and visual information leads to a further *calibration* of the *visuo-manual* system.

What the motoric does is it to set up and/or maintain specific wanted perception constellations. A perception change can have an effect on the motor skills, just as a change in the environment can also have an affect: the clearest signals are a conglomerate from intrinsic and external signals.

By connecting the environmental motor stimuli and the continuous changes in perception the organism is able to perceptive copy of effect of his motoric. The organism can become familiar with the sensory consequences of its own motoric when it becomes active under poorly stimulated environmental conditions. These stimuli, as a rule, cause changes to the situation and position of limbs, which are registered - either in a direct sensory way, and which are acknowledged by the active organism - or else indirectly registered via use of a tool. When these sensory motoric connections are established, the organism can also then correctly register environmental changes, if he himself is motor active.

These learning processes can be described as the establishment of motoric programs or as control cycles. Motor learning could then be understood as a crossing over from regulation to control. (cp. [Kalveram](#), 1981). The cerebral representation of the development of a spatial-temporal relationship of the act of carrying out movement, is however, not yet completely investigated. (cp. [Guedon et al.](#), 1998; cp. [Mai, Blaut & Hermsdörfer](#), 1993).

Tracking

Sensory motoric control cycles are experimentally operational by means of so-called [tracking-tasks](#), which have the following qualities (cp. [Hammerston](#), 1981):

- . An externally defaulted static or dynamic Input signal defines the motoric answer of the operator, which processes the task by initiation of visual and manual control processes
- . These control mechanisms generate the Output-signal
- . The difference between that output - and the input is the error value that should be held as possible from the operator

Processing these tasks requires a well matched visuo-motoric co-ordination. The predictability of the target of the INPUT signal is the most important factor for the tracking conversion. The degree of the connection between eye and hand system increases with decreasing predictability of the destination or with increasing target complexity. This result ([Xia & Barnes](#), 1999) represents a common central

sensory motoric control system of eye and hand, in the case of the tracking, which has also been suggested by other authors ([Gauthier et al.](#), 1988). On the other hand, Bock (1987, cp. [Xia & Barnes](#), 1999) accepts the existence of parallel, independent channels.

Visuo-manual System

The eye and arm are subject to different physical and dynamic qualities and consequently, have different reaction times. Furthermore, eye movements and therefore the visual recording of movement changes, are to a small extent, voluntarily controllable and may have a considerably smaller degree of freedom than the arm. While hand movement is more of a progressive type of movement, eye movement is different during saccadic and slow movements (cp. [von Cramon & Zihl](#), 1988). When it comes to spatial-temporal organization of motions, aspects of the cognitive organization and the range of automation should be taken into further considered ([Philips](#), 1986):

- . the execution of known motions after cognitive anticipation of the motions;
- . what follows unknown motions and increased demands on the behavioural reflex and
- . the execution of highly automatic patterns of movement as well as internal patterns of movement.

The participation of other subsets in the system of visual-motoric co-ordination

The adaptation of co-ordination requires an increased control over the attention processes. When it comes to the performance of visuo-motoric co-ordination, one can assume that the selective attention processes ([Bucks](#), 1997) is, during long durations of continuous acoustic activation, in a phase of an lasting attention performance. In the visual field Posner has distinguished (1987; cf. [Fimm](#), in 1997), with respect to the selective attention, three basic mechanisms, which occur after brain lesions, can also have selective damage:

- . solved attention (disengage)
- . postponed attention (move)
- . focused attention (engage).

Due to the participation of numerous fields and structures of the brain, the attention system displays a special vulnerability after every cerebral insults and dysfunction.

Furthermore, the precondition for visuo-motoric performance is an intact *visual-spatial perception*, which, among other things, includes the correct estimation of positions, distances and angles ([Kerkhoff](#), 1988, 1989).

These basic functions are highly relevant in everyday life. With spatial impairments, all practical activities can be seriously affected, which require one visual-spatial process or subsets of the spatial-constructive system.

Impairments to the motoric system

[Impairments to the motoric system](#) can occur within the framework of numerous cerebral and peripheral injuries.

As a result of an injury to the motoric cortex, central paralysis can occur in the contra-lateral half of the body. Impairments in the basal ganglion, as in the case of Huntington's chorea and morbus Parkinson, have multiple negative consequences. Impairments to the okulomotor (which can also affect the visuo-motoric co-ordination) occur within the framework of numerous neologic illnesses; in particular as result of injuries to the brain stem.

One can also observe disturbances to spatial and temporal co-ordination of the movements of extremities, as in the case of **apraxia**, which Liepmann distinguished as (1905, 1908; cf. [Goldenberg](#), 1993) ideo-motoric (incorrect imitation of gestures) and idea-toric (involuntary movement and loss of memory i.e. for the correct use of an object) apraxia. These impairments affect, in a selective manner, the planning or carrying out of motoric actions and always, in the case of unilateral lesions, the ipsilateral side the body - in relation to the lesion.

After patients have been treated in a rehabilitation clinic, they often complain of impairments to their concentration and attention. Alongside these impairments [patients](#) also complain of disturbances to their hand movement, in particular when carrying out detailed work. The deficits can affect different aspects of motoric ability, for example, the sensitivity with regard to different modalities, finger mobility, physical strength or co-ordination. The scale of the impairments can vary significantly. Some patients the lack of mobility in the hands during every day activities, others only notice it when carrying out a complicated action with the hands, for example, playing a musical instrument. Impairments to movement in the hands can only be characterized after a detailed description of these disturbances on several levels (for a full representation see - [Mai et al.](#), 1993). For a detailed diagnostic of sensory motoric impairments - see the diagnostic from Schoppe (1974; cp. [Mai et al.](#), 1993), [Platz](#) (1992), [Weber](#) (1992) and [Behreburg](#) (1992).

2.2 Aim of the training

The aim of the training is to improve the patient's [visuo-motoric co-ordination](#). In addition the training also helps to improve the patient's [level of attention](#) and reaction time.

The training works on the basis of the so-called [Tracking](#) - task. The patient has to follow a given movement (a set of lines or the movement of a pointer) on the screen by using a control panel (joystick, control dial).

The level of difficulty is established:

- . through the training tasks themselves, whereby motions always re-occur (the learning the processes encourages the anticipation of the individual patterns of movement),
- . through tasks with increasing demands on the patient's reaction ability (having to follow previously unseen pattern of movement) and
- . through stochastic movements, interruptions and reversals within a given time.

One can expect that an improvement in the [visuo-motoric co-ordination](#) will have a positive affect in the ADL field (Activities of Daily Living), as motoric disorders cause daily hindrances (eating, getting dressed, body care, writing, drawing, driving).

On the premise of maximum specificity of therapy, one should always precede the training with a differentiated disorder specific neuropsychological diagnostic (cf. [Grundlagen](#)).

Although the diagnostic protects the deficit because of the complexity of the function and therefore a precise training is complicated, which results in unsystematic practicing of precise training finger-hand-arm movements.

The following RehaCom-procedures are recommended as a support to this procedure: **Spatial operations** (RAUM), **Visuo-constructive abilities**(KONS), and **Two-dimensional operations** (VRO1).

2.3 Target groups

This procedure is recommended for training the [visuo-motoric co-ordination](#) skills of patients with impairments to their [hand control /functions](#) and [visual spatial perception](#). It is also recommended for patients who have problems with [spatial construction](#).

Defective control / functions in the hand as well as disorders in visuo-motoric co-ordination often happen as a result of injuries to the organic brain, e.g.as in the case of cerebral insults, infarcts, hemorrhages, tumours or brain trauma. However they are also known to occur after diffuse injuries to the brain. Patients suffer mostly from *paralysis, disorders in their bodily sensations (awareness of the body) or with atactic disorders in motion*. This is often the case if a patient is

suffering from pyramidal disorder or after a cerebral injury. A broad spectrum of motoric disorders occur as a result of injuries/diseases to the brain, e.g. chorea Huntington, Morbus Parkinson, multiple sclerosis or as a result of cerebral system atrophy. The following can also have a negative affect on the visuo-motoric co-ordination: deficits in the elementary [sight functions](#), disorders in the [Oculomotoric](#), disorders to the [visual-spatial](#) functions, [apraxia](#) and serious [attention disorders](#). Disorders to the visual-spatial functions and/or a dysfunctional [proprioception](#) have negative affects on the adaptive mechanisms of the continuous movement control and correction..

As a result of complex disorders to the visual-spatial deficits, patients who suffer from brain damage are primarily affected due to posterior and/or partial-occipital uni- and bilateral lesions or injuries, which in turn damage the visual system.

In particular, right-hand parietal lesions cause visual-spatial disorders or in very rare cases visual neglect itself. Patients who suffer from Hemiplegie, which affects the dominant hand, could also improve the mobility of their other hand by training with the procedure **visuo-motoric co-ordination**.

The training should be started as soon as the first signs of such disorders. The procedure can be used with children from 8 and up. Up to the age of 14 appropriate instructions for children should be used. It is also possible to use the training to the help with disorders in [attention](#)

If the patient suffers from extreme disorders to motoric functions (e.g. Ataxia), whereby the patient is not capable of using the controls; then this type of training is not recommended.

2.4 Bibliography

Backs, R.W. (1997): Psychophysiological aspects of selective and divided attention during continuous manual tracking. *Acta Psychologica*, 96, S. 167-191.

Behremburg, Ch.; Stephan, K.M. & Hömberg, V. (1992): Quantitative Erfassung von diskreten Feinmotorikstörungen bei Patienten mit linkshemisphärischem Insult. In: Mauritz, K.-H. & Hömberg V. (Hrsg.): *Neurologische Rehabilitation*. Band 2. Bern, Göttingen: Huber.

Butters, N. & Barton, M. (1970): Effect of parietal lobe damage of the performance of reversible operations in space. *Neuropsychologia*, 8, S. 205-214.

Cody, F.-W.-J.; Loevgreen, B. & Schady, W. (1993): Increased dependence upon visual information of movement performance during visuo-motor tracking in

cerebellar disorders, *Electroencephalography and Clinical Neurophysiology: Electromyography and Motor Control*, 89 (6), 399-407.

Cramon, D. Y. von (1988): Sehen. In: von Cramon, D. Y. von & Zihl, J (Hrsg.). *Neuropsychologische Rehabilitation*. Berlin: Springer-Verlag. S. 105-129.

Fimm, B. (1997): Microanalyse von Aufmerksamkeitsprozessen. In: Gauggel, S. & Kerkhoff, G. (Hrsg.): *Fallbuch der Klinischen Neuropsychologie. Praxis der Neurorehabilitation*. Göttingen: Hogrefe. S. 25-38.

Gauthier, G.M.; Vercher, J.L.; Mussa-Ivaldi, F. & Marchetti, E. (1988): Oculo-manual tracking of visual targets: Control learning, co-ordination control and co-ordination model. *Experimental Brain Research*, 73, 127-137.

Goldenberg, G. (1993): Praxie. In: Cramon, D.Y. von, Mai, N, & Ziegler, W. (Hrsg.): *Neuropsychologische Diagnostik*. Weinheim: Chapman & Hall.

Guedon, O.; Gauthier, G.; Cole, J.; Vercher, J.-L- & Blouin, J. (1998): Adaptation in Visuomanual Tracking Depends on Intact Proprioception. *Journal of Motor Behaviour*, Vol. 30 (3), S. 234-248.

Hammerton, M. (1981): Tracking. In: Holding, D.: *Human Scills*. John Wiley & Sons Ltd, S. 177-201.

Hartje, W. & Sturm, W. (1989): Räumliche Orientierungsstörungen und konstruktive Apraxie. In: Poeck, K. (Hrsg.): *Klinische Neuropsychologie*. Stuttgart, New York: Thieme Verlag.

Kalveram, K.T. (1981): Erwerb sensumotorischer Koordination unter störenden Umwelteinflüssen: Ein Beitrag zum Problem des Erlernens von Werkzeuggebrauch. *Festschrift für Heinrich Düker*, S. 336-348.

Kalveram, K.T. (1983): Programmsteuerung und Parameterverstellung bei einem sensumotorischen Lernvorgang, *Archiv für Psychologie*, Vol. 135 (3), S. 245-255.

Kerkhoff, G. (1988): Visuelle Raumwahrnehmung und Raumoperationen. In: Cramon, D. Y. von & Zihl, J (Hrsg.): *Neuropsychologische Rehabilitation*. Berlin: Springer-Verlag. S. 197-214.

Kerkhoff, G. (1989): Störungen der visuellen Raumwahrnehmung bei Patienten mit Hirnschädigung. Frankfurt am Main: Haag und Herchen Verlag.

Kerkhoff, G. & Marquardt, C. (1995): Quantitative Erfassung visuell-räumlicher Wahrnehmungsleistungen in der Neurorehabilitation. *Neurol. Rehabil.*, Vol. 2,

S. 101-106.

Kerkhoff, G., Münßinger, U. & Marquardt, C. (1993): Sehen. In: Cramon, D.Y. von, Mai, N. & Ziegler, W. (Hrsg.): Neuropsychologische Diagnostik. Weinheim: Chapman & Hall.

Kolb & Wishaw (1990): Fundamentals Of Human Neuropsychology. New York: W. H. Freeman and Company.

Mai, N. (1988). Störungen der Handfunktion. In Cramon, D. v. & Zihl, J. (Hrsg.). Neuropsychologische Rehabilitation. Berlin, Heidelberg, New York: Springer-Verlag. 360 - 385.

Mai, N.; Blaut, M. & Hermsdörfer (1993): Handfunktionen. In: Cramon, D.Y. von, Mai, N. & Ziegler, W. (Hrsg.): Neuropsychologische Diagnostik. Weinheim: Chapman & Hall.

Münßinger, U. & Kerkhoff, G. (1993): Therapie räumlich-konstruktiver und räumlich-visueller Störungen bei hirngeschädigten Patienten. Praxis Ergotherapie, Vol. 6. S. 215-221.

Neilson, P.D.; Neilson, M.D. & O'Dwyer, N.J. (1995): Adaptive optimal control of human tracking. In: Glencross, D.J. & Piek, J.P. (Eds.) (1995): Motor control and sensory motor integration: Issues and directions. Advances in psychology, 111. Amsterdam: Elsevier Science Publishers.

Philips, C.G. (1986): Movements of the Hand. In: Sherrington Lectures (7), Liverpool: University Press.

Platz, T.; Kaden, B.; Denzler, P. & Mauritz, K.-H. (1992): Zielbewegungen (Pointing) bei zentraler Parese und Apraxie - Dreidimensionale Bewegungsanalyse. In: Mauritz, K.-H. & Hömberg V. (Hrsg.): Neurologische Rehabilitation. Band 2. Bern, Göttingen: HuberVerlag.

Poeck, K. (Hrsg.) (1989): Klinische Neuropsychologie. Stuttgart, New York: Thieme Verlag.

Prosiegel, M. & Säring, W. (1988): Bewegungsfolgen. In: Cramon, D. v. & Zihl J. (Hrsg.): Neuropsychologische Rehabilitation. Berlin, Heidelberg, New York: Springer-Verlag.

Ritter, M. (1983): Diagnostik sensorischer und motorischer Funktionen. In: Groffmann, K.J. & Michel, L. (Hrsg.): Intelligenz und Leistungsdiagnostik. Göttingen, Toronto, Zürich: Hogrefe Verlag.

Vercher, J.L.; Gauthier, G.M.; Guedon, O. & Blouin, J. (1995): Self-moved target eye tracking in control and deafferented subjects: Roles of arm motor command and proprioception in arm-eye coordination, *Journal of Neurophysiology*, 76 (2), S. 1133-1144.

Vercher, J.-L.; Quaccia, D. & Gauthier, G.M. (1995): Ocular tracking of self-moved targets: Role of visual and non-visual information in visuo-oculo-manual coordination. In: Findlay, J.M. (Ed.) & Walker, R. (Ed.) (1995): *Eye Movement research: Mechanisms, processes and applications. Studies in visual information processing*, 6, Amsterdam: Elsevier Science Publishing Co., S. 99-107.

Vercher, J.-L.; Volle, M. & Gauthier, G.-M. (1993): Dynamic analysis of human visuo-oculo-manual coordination control in target tracking tasks, *Activation-, Space- and Environmental Medicine*, 64 (6), 500-506.

Weber, P., Hofmann, H., Canavan, A. & Hömberg, V. (1992): Objektivierung der Diagnostik von Hand- und Augenbewegungen. In: Mauritz, K.-H. & Hömberg, V. (Hrsg.): *Neurologische Rehabilitation. Band 2*. Bern, Göttingen: Huber Verlag.

Xia, R. & Barnes, G. (1999): Oculomanual Coordination in Tracking of Pseudorandom Target Motion Stimuli. *Journal of Motor Behaviour*, Vol. 31 (1), S. 21-38.

Yamashita, T (1988): Der Einfluß von Wahrnehmungsfaktoren auf die Leistung im Folge-Tracking, *Psychologische Beiträge*, 30 (4), S. 568-578.

Index

- A -

acoustic feedback 2, 5
 Acquisition 5
 Activities of Daily Living (ADL) 10
 adaptive training 3
 Apraxia 7, 11
 Ataxia 11
 Attention disorders 11
 Attention process 7
 Automation 7
 Awareness disorders 7

- B -

Bibliography 12

- C -

central injuries 7
 constant movement 3
 Continuous Consultation 7
 Continuous data analysis 7
 continuous tone 2
 Control mode 5
 Control parameter 5
 Co-ordination 7
 Cortexareale 11
 Cursor 1

- D -

Description of the training 1
 Disorders of the visu-omotoric co-ordination 11
 Duration of tasks 5
 Duration of training/Cons. in mins 5

- E -

Error status 2

- F -

Feedback 5
 Figure 7
 Follow-Tracking 10

- G -

Graficpool-Type 5
 Grey picture 5

- H -

Hand function disorders 11
 Hand functions 7

- I -

Input mode 5

- L -

Learning of movement 7
 Level of training 1
 limited Acquisition 5
 limitederror analysis 5

- M -

Memory 5
 mobility of fingers 7
 Motoric 7
 motoric disorders 7
 motoric programme 7
 muscel strenght 7

- N -

neuro-psychological diagnostic 10
 neuro-psychological diagnostic 11
 Number of errors 1, 2

- O -

Okulomotoric 7, 11

Orientation 5
 Outside-time 2, 7

- P -

Paralysis 7
 pattern of movement 1, 10
 Patterns of movement 3
 Pauses 7
 performance feedback 2
 performance of visuo-motoric transformation 7
 Proprioception 7

- R -

Reaction ability 10
 Reaktion time 7
 RehaCom-Procedure 10
 Repetition 3, 5
 Reproduction 5
 Rotor 1

- S -

selective attention 7
 sensory motoric control system 7
 sequential movement 3
 Sim of the training 10
 Size 5, 7
 size of the rotor 3
 spatial-constructive disorders 11
 spatial-constructive performance 7
 Specifics of the Training 11
 speed of the rotor 3
 Standard evaluation 5
 Structure of the level of difficulty 3

- T -

Target groups 11
 target orientated training 7
 Tempo 5, 7
 theoretical basics 7
 time outside 1, 3, 5
 Tracking 7
 Training parameter 5

Training time 7

- V -

visual perception 7
 visual spatial perception 7
 visuall-spatial disorders 11
 visual-spatial perception 11
 visuo-manual system 7
 visuo-motoric co-ordination 10
 voluntary movements 7